

The Traceable Calibration of Thrust Stand by Electrostatic Force

Bin Wang^{1,2}, Yong Yang³, Zhan Hu^{1,2}, Yelong Zheng³

¹Shanghai Engineering Center for Microsatellites ²Innovation Academy for Microsatellites, Chinese Academy of Sciences ³State Key Laboratory of Precision Measuring Technology and Instruments, Tianjin University

Corresponding Author

Yelong Zheng
zhengyelongby@163.com

Citation

Wang, B., Yang, Y., Hu, Z., Zheng, Y. The Traceable Calibration of Thrust Stand by Electrostatic Force. *J. Vis. Exp.* (181), e63465, doi:10.3791/63465 (2022).

Date Published

March 1, 2022

DOI

10.3791/63465

URL

jove.com/video/63465

Materials

Name	Company	Catalog Number	Comments
Motorized linear stage	Zolix	TSA50-C	Resolution 0.625 μm
Capacitance bridge	Andeen-Hagerling	AH2550A	Resolution 0.8 aF, Accuracy ± 5 PPM
High voltage source measure unit (SMU) instrument	Keithley	2410	Precision 0.012%, ± 5 μV – ± 1100 V
Laser interferometer	Renishaw	RLE10	Resolution 10 nm
Circular parallel plate capacitor	Processed by high precision grinding		The plates are processed by high precision grinding of aluminum alloy. The diameter of plate A is 6 cm, and the diameter of plate B is 4 cm.
Thrust stand	Processed by high precision grinding		Pendulum type thrust stand